



PRECISION

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.


Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean(ng/ml)	12.312	12.249	11.943	12.285	12.571	12.532
SD	0.033	0.053	0.024	0.034	0.066	0.073
CV(%)	3.948	6.382	2.911	4.106	8.103	8.956

TYPICAL DATA

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml	OD1	OD2	Average
100	0.147	0.141	0.144
50	0.292	0.262	0.277
25	0.503	0.508	0.506
12.5	0.905	0.856	0.881
6.25	1.072	1.100	1.086
3.125	1.440	1.386	1.413
1.563	1.819	1.798	1.809
0	2.218	2.273	2.246

LOD

0.649ng/ml

LINEARITY

To assess the linearity of the assay, samples were spiked with high concentrations of Goat AMBP in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

Sample	Serum(n=4)
1:300	Average % 93 Range % 83-105
1:600	Average % 95 Range % 83-106
1:1200	Average % 89 Range % 84-104
1:2400	Average % 95 Range % 82-110