



PRECISION

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.


Inter-assay Precision (Precision between assays):CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean(ng/ml)	12.310	12.249	11.953	12.475	12.560	12.532
SD	0.033	0.053	0.024	0.034	0.066	0.073
CV(%)	3.947	6.381	2.910	4.105	8.102	8.955

TYPICAL DATA

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml	OD1	OD2	Average
100	0.149	0.147	0.148
50	0.294	0.270	0.282
25	0.505	0.521	0.513
12.5	0.907	0.876	0.892
6.25	1.074	1.123	1.099
3.125	1.442	1.414	1.428
1.563	1.821	1.835	1.828
0	2.320	2.220	2.270

LOD

0.649ng/ml

LINEARITY

To assess the linearity of the assay, samples were spiked with high concentrations of Guinea Pig AMBP in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

Sample	Sample	Serum(n=4)
1:300	Average %	96
	Range %	89-108
1:600	Average %	99
	Range %	84-106



WUHAN HUAMEI BIOTECH CO., LTD.

Postal Address: Wuhan Hi-tech Medical Devices Park, Building B11,

Email: cusabio@cusabio.com

#818 Gaoxin Road, Donghu Hi-Tech Development Area, Wuhan, Hubei

Tel: +86-27-87582341

Province 430206, P.R.China

Fax: +86-27-87196150

1:1200	Average %	88
	Range %	85-109
1:2400	Average %	96
	Range %	89-118