



PRECISION

Intra-assay Precision (Precision within an assay): CV%<8%

Three samples of known concentration were tested twenty times on one plate to assess.


Inter-assay Precision (Precision between assays): CV%<10%

Three samples of known concentration were tested in twenty assays to assess.

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean(ng/ml)	12.218	12.156	11.853	12.387	12.474	12.436
SD	0.033	0.053	0.024	0.034	0.066	0.073
CV(%)	3.945	6.379	2.908	4.103	8.100	8.953

TYPICAL DATA

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



ng/ml	OD1	OD2	Average
100	0.138	0.144	0.141
50	0.282	0.266	0.274
25	0.493	0.513	0.503
12.5	0.893	0.862	0.878
6.25	1.060	1.106	1.083
3.125	1.426	1.393	1.410
1.563	1.804	1.806	1.805
0	2.202	2.283	2.243

LOD

0.649ng/ml

LINEARITY

To assess the linearity of the assay, samples were spiked with high concentrations of Sheep AMBP in various matrices and diluted with the Sample Diluent to produce samples with values within the dynamic range of the assay.

	Sample	Serum(n=4)
	1:300	Average %
Range %		88-107
1:600	Average %	96
	Range %	84-108
1:1200	Average %	86
	Range %	84-105



WUHAN HUAMEI BIOTECH CO., LTD.

Postal Address: Wuhan Hi-tech Medical Devices Park, Building B11,

Email: cusabio@cusabio.com

2818 Xiaoxin Road, Donghu Hi-Tech Development Area, Wuhan, Hubei

Tel: +86-27-87582341

430206 P.R. China

Website: www.cusabio.com www.cusabio.cn

Fax: +86-27-87196150

1:2400

Average %

95

Range %

83-112

